

# What firmware changes have taken place on the Maretron FPM100?

What firmware changes have taken place on the Maretron FPM100 pressure monitor?

Update your device using the [N2KAnalyzer](#) tool.

## **FPM100 Firmware Versions**

### **1.1.2.1 (2018-04-27)**

- In tank mode, re-enable operation range checking at 4mA (20mA report 100%).

### **1.1.1.2 (2018-04-06)**

- Change maximum operating range while operating in pressure mode to 1mA - 24mA (3681). FPM100 will report actual pressure between 4mA – 20 mA. Report minimum pressure between 1mA – 4mA. Report maximum pressure between 20mA – 24mA. Report Out-Of-Range below 1mA or above 24mA

- Always report data between 0-100% while operating in tank mode without sensor error detection (3681). FPM100 will report actual level between 4mA – 20 mA. Report 0% below 4mA. Report 100% above 20mA.

### **1.0.2.2 (2014-10-31)**

- Added support for real-time display of current readings during calibration

### **1.0.1 (2011-11-03)**

- Initial Release

Online URL:

<https://www.maretron.com/wp-content/phpkbv96/article.php?id=532>